NOTES:

1. SUBSTRATE

FUSED SILICA

2. SURFACE S2 TO BE PARALLEL WITH S1 WITHIN < 3 ARCMIN

3. COATING (APPLY ACROSS CLEAR APERTURE)

S1: Laser Mirror (266nm) R(ABS) >99.5% @ 266nm @ 45° AOI R(AVG) >99.5% @ 263 - 268nm @ 45° AOI

DAMAGE THRESHOLD,

PULSED: 2.5 J/cm² @ 266nm, 20ns, 20Hz

CW: 1MW/cm² @ 266nm

S2: NONE


4.


FINE GROUND SURFACE

5. POWER, IRREGULARITY, SURFACE QUALITY, AND COATING SPECIFICATIONS APPLY ACROSS CLEAR APERTURE

6. Rohs Compliant







SPECIFICATIONS SUBJECT TO CHANGE WITHOUT NOTICE DIMENSIONS ARE FOR REFERENCE ONLY

	S1	S2			R	Edmund Optic	S ®
SHAPE	PLANO	PLANO			w		
SURFACE FLATNESS	λ/10	N/A		ĺ		Ø12.5mm, 255 - 275nm 45°, ND:YAG LAS	SED LINIE
SURFACE QUALITY	10-5	FINE GROUND	THIRD ANGLE PROJECTION		TITLE	MIRROR	
CLEAR APERTURE	Ø11.25	N/A					CLIEFT
BEVEL	PROTECTIVE AS NEEDED	PROTECTIVE AS NEEDED	ALL DIMS IN	mm	DWG NO	47980	SHEET 1 OF 1