NOTES:

1. SUBSTRATE:
   N-BK7

2. SURFACE S2 TO BE PARALLEL TO SURFACE S1 TO WITHIN 1 ARCMIN

3. COATING (APPLY ACROSS COATING APERTURE)
   S1: NONE
   S2: NONE

4. EDGES: FINE GROUND

5. POWER, IRREGULARITY, AND SURFACE QUALITY SPECIFICATIONS APPLY
   ACROSS CLEAR APERTURE

6. TRANSMITTED WAVE FRONT OVER THE CLEAR APERTURE SHALL BE
   SPHERICAL (Y^4) -0.25λ WAVE PEAK TO VALLEY @ 587nm.
   WAVE FRONT ERROR FROM IDEAL SPHERICAL FORM SHALL BE
   LESS THEN ±0.0625 WAVES

7. ASPHERIC SURFACE DESCRIBED BY (REF. COEFFICIENT TABLE):

$$Z(Y) = \frac{\left(1/RADIUS\right)*Y^2}{1+\sqrt{1-(1+k)*\left(1/RADIUS\right)^2*Y^2}} + D*Y^2 + E*Y^4 + F*Y^6 + G*Y^8 + H*Y^{10} + J*Y^{12} + L*Y^{14}$$

8. APPLY AN ARROW POINTING TOWARDS THE ASPHEREIC SURFACE S2 WITH
   PENCIL OR PERMANENT INK

ϕ 50.00 $^{0.00}_{-0.25}$

4.00±0.20

S1   S2 6. 7.

8.   4.

COEFFIECIENT TABLE 7.

COEFFICIENT	S1	S2
k	0	0
D	0	0
E	0	1.109024E-09
G	0	0
H	0	0
J	0	0
L	0	0

	S1	S2
SHAPE	PLANO	PLANO
CLEAR APERTURE	>85	>85
SURFACE QUALITY	60-40	60-40
BEVEL	PROTECTIVE AS NEEDED	PROTECTIVE AS NEEDED

SPECIFICATIONS SUBJECT TO CHANGE WITHOUT NOTICE
DIMENSIONS ARE FOR REFERENCE ONLY

Edmund Optics®

THIRD ANGLE PROJECTION

TITLE: 50mm DIA -0.25λ ABERRATION, SPHERICAL ABERRATION PLATE

ALL DIMS IN: mm

DWG NO: 66762

SHEET 1 OF 1